

Balance Calibration Logbook

(Ref, SOP MAN-100)

Form-185 Issue Date:

MAN-100 Weighing Equipment - Checking and Calibration

Location	MODEL	ID. NO.	STD Wt.1	TOLERANCE	STD Wt.2	TOLERANCE	STD Wt.3	TOLERANCE
E.G. Tablet Plant	PG203-S	Various	0.5 g	0.495 to 0.505 g	10 g	9.9 to 10.1 g	200 g	198 to 202 g